



Certificate of Analysis

Product: COMPLEMENT C4
Catalogue Number: BXC0861A
Batch Number: 171431
Manufactured Date: 2017/06
Expiry: 2019/03

This product has been tested by Quality Control and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

The following results were obtained:

Control Description	Lot Number	Target (mg/dl)	Range (mg/dl)	Result (mg/dl)
Protein Control L1	641-01-018	16.2	12.96-19.41	15.52
Protein Control L2	642-02-020	27	21.6-32.35	25.015

QC PASS/FAIL: **PASS**

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Cherith Morris* (QC Manager)

Date: 2017/06