



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **CALCIUM** MANUFACTURE DATE: 04/13/2010  
 CATALOG #: C503-480  
 LOT #: 25133 EXP. DATE: 11/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
CALCIUM BUFFER REAGENT	C503-A	21264	11/2012
CALCIUM COLOR REAGENT	C503-B	21385	11/2012
CALCIUM STANDARD (10 mg/dl)	C503-C	21386	11/2012

BLANK: 0.88 AT 550 NM

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-Matrix)	<u>TARGET</u> (mg/dl)
15271	03/2012	7.2 – 8.8	8.0
15272	03/2012	10 – 12	11

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	7.9	7.7	7.8	2/2
15272	11.3	11.1	11.2	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED  
BY:

Quality Control

DATE: 04/13/2010