TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME:	**CALCIUM**	MANUFACTURE DATE:	12/16/2014
CATALOG #:	C503-480		
LOT #:	58546	EXP. DATE:	06/2017

COMPONENTS:	CATALOG #:	LOT#:	EXP. DATE:
CALCIUM BUFFER REAGENT	C503-A	58295	12/2017
CALCIUM COLOR REAGENT	C503-B	58296	12/2017
CALCIUM STANDARD (10 mg/dl)	C503-C	55160	06/2017

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE	TARGET
		(TC-Matrix)	(mg/dl)
27621	12/2015	7.5 – 11.2	9.3
27602	03/2015	11.9 – 14.6	13.3

TEST VALUES/RESULTS

SERA ID	MEAN	PASS/FAIL
27621	10.1	PASS
27602	14.2	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Jacqueline Sanchez
Quality Control

DATE: 12/17/2014