



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CHLORIDE** MANUFACTURE DATE: 11/17/2010
 CATALOG #: C501-480
 LOT #: 28383 EXP. DATE: 11/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
CHLORIDE REAGENT	C501-A	28452	10/2013
CHLORIDE STANDARD (100 mEq/L)	C501-B	22249	11/2012

BLANK: 0.064 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mEq/L)
15271	03/2012	77.2-110.9	99
15272	03/2012	78-102	90

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u>#PASS</u>
15271	85.6	87.3	86.45	2/2
15272	90.5	94.0	92.25	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Bassey* DATE: 11/18/2010
Quality Control