



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CHOLESTEROL LIQUID** MANUFACTURE DATE: 12/02/2010
 CATALOG #: C507-480
 LOT #: 29064 EXP. DATE: 05/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. CHOLESTEROL LIQUID REAGENT	C507-A	29067	05/2012
2. CHOLESTEROL STANDARD (200 mg/dL)	C509-B	29008	11/2013

BLANK: 0.000 AT 505 NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dL)
15271	03/2012	221(245.5)270	245.5
15272	03/2012	90.7(100.8)110.9	100.8

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	243.3	244.5	243.9	2/2
15272	95.2	94.3	94.75	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Casan* DATE: 12/02/2010
 Quality Control