



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CHOLESTEROL LIQUID** MANUFACTURE DATE: 08/31/2011
 CATALOG #: C507-480
 LOT #: 34642 EXP. DATE: 02/2013

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. CHOLESTEROL LIQUID REAGENT	C507-A	34626	02/2013
2. CHOLESTEROL STANDARD (200 mg/dL)	C509-B	33330	07/2014

BLANK: 0.017 AT 505 NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
15281	09/2012	201 - 314	258
15282	09/2012	85 - 130	108

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	254	254	254	2/2
15282	105	105	105	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 08/31/2011
Quality Control