TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME:	**CHOLESTEROL - 10**	MANUFACTURE DATE:	7/21/2009
CATALOG #:	C509-400		
LOT #:	20123	EXP. DATE:	11/2011

COMPONENTS:	1.CHOLESTEROL REAGENT	LOT #:	17432	EXP. DATE:	11/2011
	2.CHOLESTEROL STANDARD	LOT #:	19613	EXP. DATE:	06/2012

BLANK: 0.010 AT 505 NM

STANDARD: 200 mg/dL

TOLERANCE LEVELS OF SERUM CONTROL

SERA ID	EXP. DATE	TOLERANCE RANGE (On TC-84)	TARGET (mg/dL)
15261	08/2011	226 – 276	251
15262	08/2011	91 – 123	107

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15261	256.8	253.1	255.0	2/2
15262	110.7	109.6	110.2	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control

DATE: 7/21/2009