



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CHOLESTEROL - 10** MANUFACTURE DATE: 05/14/2010
 CATALOG #: C509-400
 LOT #: 24706 EXP. DATE: 12/2022

COMPONENTS: 1.CHOLESTEROL REAGENT LOT #: 23771 EXP. DATE: 12/2012
 2.CHOLESTEROL STANDARD LOT #: 23346 EXP. DATE: 12/2012

BLANK: 0.010 AT 505 NM

STANDARD: 200 mg/dL


TOLERANCE LEVELS OF SERUM CONTROL

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dL)
15271	03/2012	210 – 258	234
15272	03/2012	85.8 – 104.8	95.3

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	242.7	240.2	241.5	2/2
15272	94.0	97.4	95.7	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE D BY:	 Quality Control	DATE :	<u>05/14/2010</u>
------------------	--	-----------	-------------------