

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **CHOLESTEROL - 10**      MANUFACTURE DATE: 07/09/2010  
CATALOG #: C509-400  
LOT #: 26604      EXP. DATE: 04/2013

COMPONENTS: 1.CHOLESTEROL REAGENT      LOT #: 25353      EXP. DATE: 05/2013  
2.CHOLESTEROL STANDARD      LOT #: 25174      EXP. DATE: 04/2013

BLANK: 0.009      AT 505 NM

STANDARD: 200 mg/dL

**TOLERANCE LEVELS OF SERUM CONTROL**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dL)
15271	03/2012	210 - 258	234
15272	03/2012	85.8 - 104.8	95.3

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	254.6	249.7	252.2	2/2
15272	93.1	90.2	92.0	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE D BY:	 Quality Control	DATE :	<u>07/09/2010</u>
------------------	--	-----------	-------------------