



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CHOLINESTERASE** MANUFACTURE DATE: 05/13/2010
CATALOG #: C511-60
LOT #: 25108 EXP. DATE: 06/2011

COMPONENT: CHOLINESTERASE REAGENT LOT#: 18421 EXP. DATE: 06/2011

REAGENT BLANK VS. WATER: 0.249 AT 405 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (IU/L)</u>
15271	03/2012	1448-1767	1607
15272	03/2012	1751-2140	1946

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	1494	1505	1500	2/2
15272	1935	2043	1989	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: 
Quality Control DATE: 05/14/2010