



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CHOLINESTERASE** MANUFACTURE DATE: 12/20/2010
CATALOG #: C511-60
LOT #: 29686 EXP. DATE: 08/2012

COMPONENT: CHOLINESTERASE REAGENT LOT#: 25481 EXP. DATE: 08/2012

REAGENT BLANK VS. WATER: 0.321 AT 405 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(IU/L)</u>
15271	03/2012	1448-1767	1607
15272	03/2012	1751-2140	1946

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	1532	1535	1533.5	2/2
15272	1765	1808	1786.5	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: 
Quality Control DATE: 12/20/2010