

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CHOLINESTERASE** MANUFACTURE DATE: 01/15/2015
CATALOG #: C511-60
LOT #: 58874 EXP. DATE: 05/2017

COMPONENT: CHOLINESTERASE REAGENT LOT#: 58633 EXP. DATE: 05/2017

TOLERANCE LEVELS OF SERUM CONTROLS


<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (IU/L)
27621	12/2015	1886 – 3503	2645
27602	03/2015	878 – 1318	1098

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27621	2478	PASS
27602	1005	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


Jacqueline Sanchez
Quality Control

DATE:

01/15/2015