



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CK – NAC (KINETIC)** MANUFACTURE DATE: 10/23/2009
 LOT #: 21122 EXP. DATE: 06/2011

COMPONENT: CK-NAC REAGENT LOT#: 17431 EXP. DATE: 06/2011

REAGENT BLANK VS. WATER: 0.347 AT 340 NM


TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (U/L)</u>
15271	03/2012	105.5 – 158.2	131.9
15272	03/2012	371.9 – 557.9	464.9

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	130.5	141.6	136.1	2/2
15272	446.8	495.0	471	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE D BY:	 Quality Control	DATE :	<u>10/23/2009</u>
------------------	--	-----------	-------------------