



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **CK - NAC (KINETIC)**                      MANUFACTURE DATE: 05/20/2010  
 CATALOG #: C512-60  
 LOT #: 25770    EXP. DATE: 03/2012

COMPONENT: CK-NAC REAGENT (CAT#: C512-A)      LOT#: 20620      EXP. DATE: 03/2012

REAGENT BLANK VS. WATER: 0.413      AT 340 NM

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (IU/L)</u>
15271	03/2012	105.5 - 158.2	131.9
15272	03/2012	371.9 - 557.9	464.9

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	145.3	134.2	139.8	2/2
15272	483.9	476.1	480	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE      *[Signature]*      DATE      05/20/2010  
 D BY:      Quality Control      :