**TECO DIAGNOSTICS**

**1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.**
**Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880**
**Fax: (714) 463-1169**
**www.tecodiag.com**

# CERTIFICATE OF ANALYSIS
## (FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CK – NAC (KINETIC)** MANUFACTURE DATE: 06/14/2010

LOT #: 26161 EXP. DATE: 08/2012

COMPONENT: CK-NAC REAGENT LOT#: 23363 EXP. DATE: 08/2012

REAGENT BLANK VS. WATER: 0.289 AT 340 NM

### TOLERANCE LEVELS OF SERUM CONTROLS

| SERA ID | EXP. DATE | TOLERANCE RANGE | TARGET (IU/L) |
|---|---|---|---|
| 15271 | 03/2012 | 105.5 – 158.2 | 131.9 |
| 15272 | 03/2012 | 371.9 – 557.9 | 464.9 |

### TEST VALUES/RESULTS

| SERA ID | 1 | 2 | MEAN | # PASS |
|---|---|---|---|---|
| 15271 | 141.2 | 131.3 | 136.3 | 2/2 |
| 15272 | 470.7 | 495.8 | 483.3 | 2/2 |

COMMENTS: VISUAL CHECK OK.

| APPROVED BY: | Quality Control | DATE: | 05/14/2010 |
|---|---|---|---|