TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CK – NAC (KINETIC)** MANUFACTURE DATE: 01/24/2011
CATALOG #: C512-60
LOT #: 28388 EXP. DATE: 08/2012

COMPONENT: CK-NAC REAGENT (CAT#: C512-A) LOT#: 23363 EXP. DATE: 08/2012

REAGENT BLANK VS. WATER: 0.298 AT 340 NM

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE	TARGET (IU/L)
15271	03/2012	110.2-165.4	137.8
15272	03/2012	386.3-579.5	482.9

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15271	138	139	138.5	2/2
15272	521	528	524.5	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 01/24/2011