TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CK – NAC (KINETIC)** MANUFACTURE DATE: 05/07/2014
CATALOG #: C512-60
LOT #: 54621 EXP. DATE: 10/2016

COMPONENT: CK-NAC REAGENT (CAT#: C512-A) LOT#: 53924 EXP. DATE: 10/2016

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE (On TC-Matrix)	TARGET (IU/L)
27611	05/2015	103.4 – 155.2	129.3
27602	03/2015	353.5 – 530.2	441.8

TEST VALUES/RESULTS

SERA ID	MEAN	PASS/FAIL
27611	144.7	PASS
27602	487.3	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Jacqueline Sanchez
Quality Control

DATE: 05/08/2016