TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CK-MB REAGENT SET** MANUFACTURE DATE: 9/30/2011
CATALOG #: C614-30
LOT #: 20866 EXP. DATE: 5/2011

COMPONENTS:					
1. CK-MB REAGENT	LOT#:	17431	EXP. DATE:	6/2011
2. CK-MB DILUENT	LOT#:	19292	EXP. DATE:	5/2011

TOLERANCE LEVELS OF SERUM CONTROLS

CONTROLS	LOT#	EXP. DATE	TOLERANCE RANGE (TC MATRIX)	TARGET (IU/L)
LEVEL I	I-68-31	12/2010	<25	
LEVEL II	I-68-29	12/2010	38.7 – 58.1	48.4
LEVEL III	I-68-30	12/2010	72.9 – 109.4	91.2

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
LEVEL I	10	10	10	2/2
LEVEL II	45	46	46	2/2
LEVEL III	83	85	84	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: QA/QC DATE: 9/30/2009