



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **CK-MB REAGENT SET**      MANUFACTURE DATE: 02/04/2011  
 CATALOG #: C614-60  
 LOT #: 29433      EXP. DATE: 08/2012

COMPONENTS:    1. CK-MB REAGENT      LOT#: 23363    EXP. DATE: 08/2012  
                          2. CK-MB DILUENT      LOT#: 30510    EXP. DATE: 02/2013

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>CONTROLS</u>	<u>LOT#</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (TC MATRIX)	<u>TARGET</u> (IU/L)
LEVEL I	I-74-25	12/2012	<25	17
LEVEL II	I-74-26	12/2012	40-60	43
LEVEL III	I-74-27	12/2012	73-105	79

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
LEVEL I	2	7	4.5	2/2
LEVEL II	56	53	54.5	2/2
LEVEL III	96	96	96	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. [Signature]*      DATE: 02/07/2011  
 BY: Quality Control