



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: CK-MB REAGENT SET MANUFACTURE DATE: 06/25/2015
CATALOG #: C614-30
LOT #: 61538 EXP. DATE: 02/2017

COMPONENTS: 1. CK-MB REAGENT LOT#: 56123 EXP. DATE: 02/2017
 2. CK-MB DILUENT LOT#: 60357 EXP. DATE: 06/2017

TOLERANCE LEVELS OF SERUM CONTROLS

<u>CONTROLS</u>	<u>LOT#</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (I/U/L)
LEVEL I	1-78-24	02/2016	<25	N/A
LEVEL II	1-78-25	02/2016	40 - 60	50

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
LEVEL I	24	Pass
LEVEL II	46	Pass

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


Jacqueline Sanchez
Quality Control

DATE:

06/25/2015