



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CREATININE (KINETIC)** MANUFACTURE DATE: 12/10/2010
 CATALOG #: C514-480
 LOT #: 29368 EXP. DATE: 03/2013

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT #:</u>	<u>EXP. DATE:</u>
1. PICRIC ACID REAGENT	C514-A	29191	11/2013
2. SODIUM HYDROXIDE (NaOH) REAGENT	C514-B	29609	11/2013
3. STANDARD (5.0 mg/dL)	C514-C	28098	03/2013

BLANK: 0.108 505

TOLERANCE LEVELS OF SERUM CONTROL

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dL)
15271	03/2012	0.97-2.28	1.63
15272	03/2012	3.91-7.28	5.59

TEST VALUES/DATA

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	1.9	1.9	1.9	2/2
15272	6.1	5.8	6.0	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 12/10/2010
Quality Control