CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CREATININE (KINETIC)**
CATALOG #: C514-480
LOT #: 36518

MANUFACTURE DATE: 12/08/2011
EXP. DATE: 05/2014

COMPONENTS:	CATALOG #:	LOT #:	EXP. DATE:
1. PICRIC ACID REAGENT	C514-A	36267	11/2014
2. SODIUM HYDROXIDE (NaOH) REAGENT	C514-B	36268	11/2014
3. STANDARD (5.0 mg/dL)	C514-C	36269	05/2014

BLANK: 0.146 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROL

SERA ID	EXP. DATE	TOLERANCE RANGE	TARGET (mg/dL)
15281	09/2012	1.3 – 2.1	1.7
15282	09/2012	4.8 – 7.1	5.9

TEST VALUES/DATA

SERA ID	1	2	MEAN	# PASS
15281	1.8	1.7	1.8	2/2
15282	6.2	6.1	6.2	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: _____ DATE: 12/13/2011

Nour Qasqas
Quality Control Lab. Director