



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **CREATININE (KINETIC)** MANUFACTURE DATE: 05/20/2014
CATALOG #: C514-480
LOT #: 55017 EXP. DATE: 04/2016

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT #:</u>	<u>EXP. DATE:</u>
1. PICRIC ACID REAGENT	C514-A	52774	02/2017
2. SODIUM HYDROXIDE (NaOH) REAGENT	C514-B	52775	02/2017
3. STANDARD (5.0 mg/dL)	C514-C	50853	04/2016

TOLERANCE LEVELS OF SERUM CONTROL


<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
27611	05/2015	0.9-1.3	1.1
27602	03/2015	4.2-6.4	5.3

TEST VALUES/DATA

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	1.1	PASS
27602	5.5	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


Jacqueline Sanchez
Quality Control

DATE:

05/21/2014