**TECO DIAGNOSTICS**

**1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.**
**Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880**
**Fax: (714) 463-1169**
**www.tecodiag.com**

# CERTIFICATE OF ANALYSIS
# FINISHED KIT QUALITY CONTROL

PRODUCT NAME: **BILIRUBIN - DIRECT** MANUFACTURE DATE: 01/16/2009
CATALOG #: B538-480
LOT #: 17962 EXP. DATE: 01/2012

| COMPONENTS: | | | | | |
|---|---|---|---|---|---|
| 1. DIRECT BILIRUBIN REAGENT | LOT#: | 17572 | EXP. DATE: | 01/2012 |
| 2. NITRITE REAGENT | LOT#: | 17491 | EXP. DATE: | 01/2012 |
| 3. CALIBRATOR (5 mg/dL) | LOT#: | 17492 | EXP. DATE: | 01/2012 |

BLANK: 0.006 AT 550 NM

## TOLERANCE LEVELS OF SERUM CONTROLS

| SERA ID | EXP. DATE | TOLERANCE RANGE (On TC-84) | TARGET (mg/dl) |
|---|---|---|---|
| 15221 | 08/2009 | 0.5 – 0.7 | 0.6 |
| 15222 | 08/2009 | 1.6 – 2.4 | 2.0 |

## TEST VALUES/RESULTS

| SERA ID | 1 | 2 | MEAN | # PASS |
|---|---|---|---|---|
| 15221 | 0.6 | 0.5 | 0.55 | 2/2 |
| 15222 | 1.8 | 1.8 | 1.8 | 2/2 |

COMMENTS: VISUAL CHECK OK.

APPROVED BY: QA/QC DATE: 01/16/2009