



CERTIFICATE OF ANALYSIS
FINISHED KIT QUALITY CONTROL

PRODUCT NAME: **BILIRUBIN - DIRECT** MANUFACTURE DATE: 7/30/2009
CATALOG #: B538-480
LOT #: 20047 EXP. DATE: 01/2012

COMPONENTS: 1. DIRECT BILIRUBIN REAGENT LOT#: 17940 EXP. DATE: 01/2012
2. NITRITE REAGENT LOT#: 19781 EXP. DATE: 06/2012
3. CALIBRATOR (5 mg/dL) LOT#: 19780 EXP. DATE: 06/2012

BLANK: 0.003 AT 550 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dl)
15261	08/2011	0.2 - 0.6	0.4
15262	08/2011	1.5-2.3	1.9

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15261	0.6	0.6	0.6	2/2
15262	1.6	1.6	1.6	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:

Quality Control

DATE: 7/30/2009