TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
FINISHED KIT QUALITY CONTROL

PRODUCT NAME:	**BILIRUBIN - DIRECT**	MANUFACTURE DATE:	9/7/2010
CATALOG #:	B538-480		
LOT #:	27405	EXP. DATE:	06/2012

COMPONENTS:	1. DIRECT BILIRUBIN REAGENT	LOT#:	25048	EXP. DATE:	04/2013
	2. NITRITE REAGENT	LOT#:	27013	EXP. DATE:	08/2013
	3. CALIBRATOR (5 mg/dL)	LOT#:	19780	EXP. DATE:	06/2012

BLANK: 0.010 AT 550 NM

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE (On TC-84)	TARGET (mg/dl)
15271	03/2012	0.0-0.8	0.4
15272	03/2012	1.2-2.5	1.9

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15271	0.2	0.5	0.35	2/2
15272	2.1	2.0	2.05	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 9/7/2010