

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **GAMMA-GT (SOLUBLE)** MANUFACTURE DATE : 01/14/2015  
CATALOG #: G571-150  
LOT#: 58829 EXP. DATE : 06/2017

COMPONENTS: GGT SOLUBLE LOT# 57928 EXP. DATE: 06/2017

**TOLERANCE LEVELS OF SERUM CONTROL**


| <u>SERA ID</u> | <u>EXP. DATE</u> | <u>TOLERANCE RANGE</u><br>(on TC-Matrix) | <u>TARGET</u><br>(IU/L) |
|----------------|------------------|--|-------------------------|
| 27621          | 12/2015          | 30.2 – 45.4                              | 37.8                    |
| 27602          | 03/2015          | 109.2 – 163.8                            | 136.5                   |

**TEST DATA**

| <u>SERA ID</u> | <u>MEAN</u> | <u>PASS/FAIL</u> |
|----------------|-------------|------------------|
| 27621          | 37.6        | PASS             |
| 27602          | 129.1       | PASS             |

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

  
Jacqueline Sanchez  
Quality Control

DATE:

