



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **GLUCOSE (OXIDASE) LIQUID** MANUFACTURE DATE: 06/22/2010
 CATALOG #: G520-480
 LOT #: 25678 EXP. DATE: 11/2011

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT #:</u>	<u>EXP. DATE:</u>
1. GLUCOSE (OXIDASE) LIQUID REAGENT	G520-A	25817	11/2011
2. GLUCOSE STANDARD (100 mg/dL)	G518-B	24657	03/2012

REAGENT BLANK VS. WATER: 0.024 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(mg/dL)</u>
15271	03/2012	69-105	87
15272	03/2012	223-335	279

TEST VALUES / RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	81.2	94.4	87.8	2/2
15272	274.2	285.4	279.8	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Bassey* DATE: 06/22/2010
Quality Control