



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **GLUCOSE OXIDASE**      MANUFACTURE DATE: 12/16/2009  
 CATALOG #: G519-1L  
 LOT #: 22992      EXP. DATE: 05/2012

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT #</u>	<u>EXP. DATE</u>
GLUCOSE OXIDASE REAGENT	G519-A	19420	05/2012
GLUCOSE STANDARD (100 mg/dL)	G518-B	21637	11/2012

BLANK: 0.015 AT 505 NM

**TOLERANCE LEVELS OF CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
15271	03/2012	71 – 107	89
15272	03/2012	237 – 355	296

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	94.2	84.1	89	2/2
15272	263.5	269.9	266.7	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

  
Quality Control

DATE: 12/16/2009