

## TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
 Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
 Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **GLUCOSE OXIDASE** MANUFACTURE DATE: 06/24/2010  
 CATALOG #: G519-1L  
 LOT #: 26281 EXP. DATE: 03/2012

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT #</u>	<u>EXP. DATE</u>
GLUCOSE OXIDASE REAGENT	G519-A	19420	05/2012
GLUCOSE STANDARD (100 mg/dL)	G518-B	24657	03/2012

BLANK: 0.000 AT 505 NM


**TOLERANCE LEVELS OF CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
15271	03/2012	71 - 107	89
15272	03/2012	237 - 355	296

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	80.8	81.6	81.2	2/2
15272	247.8	241.4	244.6	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE D BY:	 Quality Control	DATE	06/24/2010
		:	