



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **GLUCOSE OXIDASE** MANUFACTURE DATE: 01/10/2011
 CATALOG #: G519-1L
 LOT #: 29961 EXP. DATE: 03/2012

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT #</u>	<u>EXP. DATE</u>
GLUCOSE OXIDASE REAGENT	G519-A	25540	06/2013
GLUCOSE STANDARD (100 mg/dL)	G518-B	24657	03/2012

BLANK: 0.016 AT 505 NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
15271	03/2012	71-107	89
15272	03/2012	237-355	296

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	88.7	87.3	88	2/2
15272	272.4	277.9	275.15	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 01/10/2011