



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **GLUCOSE OXIDASE** MANUFACTURE DATE: 04/25/2011
CATALOG #: G519-1L
LOT #: 32187 EXP. DATE: 06/2013

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT #</u>	<u>EXP. DATE</u>
GLUCOSE OXIDASE REAGENT	G519-A	25540	06/2013
GLUCOSE STANDARD (100 mg/dL)	G518-B	29722	11/2013

BLANK: -0.003 AT 505 NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
15281	09/2012	71-106	89
15272	03/2012	237-355	296

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	81.8	82.4	82.1	2/2
15272	274.3	274.6	274.45	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE 04/26/2011

: