



CERTIFICATE OF ANALYSIS
(BATCH RELEASE)

RESIN LOT #: 090211

LYSING REAGENT LOT #: 090311

SAMPLE	LOT #	OBSERVED VALUE	EXPECTED VALUE
REFERENCE I	052541	10.0	10.0
REFERENCE II	090981	10.0	10.0
STANDARD	071311	9.8/9.9	10.0±0.2
STANDARD	072011	9.8/9.9	10.0±0.2
NORMAL CONTROL	071711	7.0/7.0	7.0±0.6
NORMAL CONTROL	101082	6.4/6.3	6.3±0.6
ELEVATED CONTROL	071712	14.3/14.4	14.5±0.9
ELEVATED CONTROL	102082	20.0/20.7	20.1±0.9
REFERENCE I			
REFERENCE II			

RESIN VOL = 38%

9/4/11

DATE

HWA

[Signature]
 QUALITY ASSURANCE