



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **GLYCOHEMOGLOBIN**      MANUFACTURE DATE: 04/21/2011  
 CATALOG #: G540-40  
 LOT #: 32111      EXP. DATE: 03/2013

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT#</u>	<u>EXP. DATE</u>
1. RESIN REAGENT	G540-A	40411	04/2013
2. LYSING REAGENT	G540-B	32511	03/2013
3. STANDARD (10%)	G540-C	11811	01/2014

**CONTROL TOLERANCE LEVELS**

<u>CONTROL OR STANDARD</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (%)</u>
Standard	11811	01/2014	8.5 – 11.5	10.0
Normal	62001	06/2013	5.9 – 8.1	7.0
Elevated	13002	01/2013	11.9 – 16.1	14.0

**TEST VALUES/RESULTS**

<u>CONTROL OR STANDARD</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
Standard	9.9	10.1	10.0	2/2
Normal	6.3	6.5	6.4	2/2
Elevated	13.7	13.9	13.8	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE D BY:	 Quality Control	DATE :	<u>04/21/2011</u>
------------------	--	-----------	-------------------