



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **GLYCOHEMOGLOBIN** MANUFACTURE DATE: 09/14/2010
 CATALOG #: G540-40
 LOT #: 27522 EXP. DATE: 03/2012

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT#</u>	<u>EXP. DATE</u>
1. RESIN REAGENT	G540-A	080701	08/2012
2. LYSING REAGENT	G540-B	31101	03/2013
3. STANDARD (10%)	G540-C	050501	05/2012

CONTROL TOLERANCE LEVELS

<u>CONTROL OR STANDARD</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (%)</u>
Standard	050501	02/2013	8.5 – 11.5	10.0
Normal	102091	10/2012	5.9 – 8.1	7.0
Elevated	102092	10/2012	11.9 – 16.1	14.0

TEST VALUES/RESULTS

<u>CONTROL OR STANDARD</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
Standard	10.1	9.9	10.0	2/2
Normal	7.5	7.3	7.4	2/2
Elevated	11.9	14.4	13.1	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE D BY:	 Quality Control	DATE :	<u>09/14/2010</u>
------------------	--	-----------	-------------------