



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **GLYCOHEMOGLOBIN** MANUFACTURE DATE: 11/23/2010
 CATALOG #: G540-40
 LOT #: 29200 EXP. DATE: 08/2012

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT#</u>	<u>EXP. DATE</u>
1. RESIN REAGENT	G540-A	110901	11/2012
2. LYSING REAGENT	G540-B	80801	08/2012
3. STANDARD (10%)	G540-C	101201	10/2013

CONTROL TOLERANCE LEVELS

<u>CONTROL OR STANDARD</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (%)</u>
Standard	101201	10/2013	8.5 – 11.5	10.0
Normal	13001	01/2013	5.9 – 8.1	7.0
Elevated	102092	10/2012	11.9 – 16.1	14.0

TEST VALUES/RESULTS

<u>CONTROL OR STANDARD</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
Standard	9.2	10.8	10	2/2
Normal	8.1	7.7	7.9	2/2
Elevated	13	12.7	12.85	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE D BY:	 Quality Control	DATE :	<u>11/23/2010</u>
------------------	--	-----------	-------------------