

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **GLYCOHEMOGLOBIN** MANUFACTURE DATE: 12/07/2011
CATALOG #: G540-40
LOT #: 36552 EXP. DATE: 10/2013

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT#</u>	<u>EXP. DATE</u>
1. RESIN REAGENT	G540-A	100511	10/2013
2. LYSING REAGENT	G540-B	100611	10/2013
3. STANDARD (10%)	G540-C	151411	05/2014

CONTROL TOLERANCE LEVELS

<u>CONTROL OR STANDARD</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (%)</u>
Standard	072011	07/2014	8.5 – 11.5	10.0
Normal	071711	07/2014	5.9 – 8.1	7.0
Elevated	071712	07/2014	11.9 – 16.1	14.0

TEST VALUES/RESULTS

<u>CONTROL OR STANDARD</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
Standard	10.1	9.9	10	2/2
Normal	6.8	6.3	6.5	2/2
Elevated	14.5	15.1	14.8	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:



DATE 12/08/2011

:

Nour Qasqas
Director of Quality Control