



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **GLYCOHEMOGLOBIN**      MANUFACTURE DATE: 08/21/2014  
 CATALOG #: G540-40  
 LOT #: 56741      EXP. DATE: 07/2016

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT#</u>	<u>EXP. DATE</u>
1. RESIN REAGENT	G540-A	071741	07/2016
2. LYSING REAGENT	G540-B	080741	08/2016
3. STANDARD (10%)	G540-C	040941	04/2017

**CONTROL TOLERANCE LEVELS**


<u>CONTROL</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC-84)	<u>TARGET</u> (%)
Normal	060731	06/2016	6.1 – 7.3	6.7
Elevated	100522	10/2015	12.8 – 14.6	13.7

**TEST VALUES/RESULTS**

<u>CONTROL</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
Normal	6.8	PASS
Elevated	13.7	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

  
 Jacqueline Sanchez  
 Quality Control

DATE:

8/22/2014