



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **GLYCOHEMOGLOBIN** MANUFACTURE DATE: 04/29//2015
 CATALOG #: G540-40
 LOT #: 60734 EXP. DATE: 04/2017

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT#</u>	<u>EXP. DATE</u>
1. RESIN REAGENT	G540-A	041851	04/2017
2. LYSING REAGENT	G540-B	040351	04/2017
3. STANDARD (10%)	G540-C	041651	04/2018

CONTROL TOLERANCE LEVELS


<u>CONTROL</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u>
Normal	060731	06/2016	(on TC-84) 6.1 – 7.3	(%) 6.7
Elevated	1005222	10/2015	12.8 – 14.6	13.7

TEST VALUES/RESULTS

<u>CONTROL</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
Normal	7.0	PASS
Elevated	14.3	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


 Jacqueline Sanchez
 Quality Assurance Manager

DATE:

04/29/2015