



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **HBDH** MANUFACTURE DATE: 9/9/2009
 CATALOG#: H620-60
 LOT #: 27541 EXP. DATE: 09/2012

COMPONENT: HBDH LOT#: 27376 EXP. DATE: 09/2012

BLANK: 0.018 AT 340 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u>
15271	03/2012	100.4 – 135.8	118.1
15272	03/2012	256.4 – 346.9	301.6

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	108.1	108.1	108.1	2/2
15272	264.8	277.6	271.2	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Pasquini* DATE: 09/14/2010
Quality Control