



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **HDL – CHOLESTEROL (PEG)**      MANUFACTURE DATE: 12/11/2009  
 CATALOG #: H511-20  
 LOT #: 22988      EXP. DATE: 5/2012

COMPONENTS: HDL CHOLESTEROL REAGENT(PEG)    LOT #: 21579    EXP. DATE: 05/2012  
 HDL CHOLESTEROL STANDARD    LOT #: 21580    EXP. DATE: 05/2012

BLANK: 0.020    AT 505 NM

STANDARD: 50 mg/dL

**TOLERANCE LEVELS OF SERUM CONTROL**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(mg/dL)</u>
910701 I	7/2011	26 – 40	33
910701 II	7/2011	99 – 135	117

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
910701 I	34.6	38.9	36.8	2/2
910701 II	119.9	129.6	124.6	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Pasqua*      DATE: 12/11/2009  
 Quality Control