



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **HDL – CHOLESTEROL (PEG)** MANUFACTURE DATE: 07/21/2010
 CATALOG #: H511-20
 LOT #: 26735 EXP. DATE: 08/2012

COMPONENTS: HDL CHOLESTEROL REAGENT(PEG) LOT #: 26459 EXP. DATE: 12/2012
 HDL CHOLESTEROL STANDARD LOT #: 24650 EXP. DATE: 08/2012

BLANK: -0.031 AT 505 NM

STANDARD: 50 mg/dL

TOLERANCE LEVELS OF SERUM CONTROL

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(mg/dL)</u>
910701 I	7/2011	26-40	35
910701 II	7/2011	99-135	117

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
910701 I	38.1	34.9	36.5	2/2
910701 II	115.0	119.0	117.0	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 07/21/2010