



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **HDL – CHOLESTEROL (PEG)** MANUFACTURE DATE: 03/17/2015
 CATALOG #: H511-20
 LOT #: 60034 EXP. DATE: 02/2017

COMPONENTS: HDL CHOLESTEROL REAGENT(PEG) LOT #: 56601 EXP. DATE: 02/2017
 HDL CHOLESTEROL STANDARD LOT #: 60027 EXP. DATE: 09/2017

TOLERANCE LEVELS OF SERUM CONTROL

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(mg/dL)</u>
407302-073	02/2017	14 - 22	18
324001-240	11/2016	42 - 58	50

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
407302-073	19.0	PASS
324001-240	51.3	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Jenifer Ohta
 Jenifer Ohta
 Quality Assurance Manager

DATE: 03/17/2015