**TECO DIAGNOSTICS**

**1268 N. LAKEVIEW AVE., ANAHEIM, CA 92807, U.S.A.**
**PHONE: (714) 463-1122 or (800) 222-9880**
**FAX: (714) 463-1169**
**www.tecodiag.com**

# CERTIFICATE OF ANALYSIS
# (FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **HEMOGLOBIN** MANUFACTURE DATE: 03/17/2010
CATALOG #: H526-6L
LOT #: 24494 EXP. DATE: 06/2012

| | | | | | |
|---|---|---|---|---|---|
| COMPONENTS: | HEMOGLOBIN REAGENT | LOT#: | 19708 | EXP. DATE: | 06/2012 |
| | HEMOGLOBIN STANDARD | LOT#: | 19710 | EXP. DATE: | 06/2012 |

BLANK: 0.005 AT 550 NM

## TOLERANCE LEVELS OF SERUM CONTROLS

| SERA ID | LOT # | EXP. DATE | TOLERANCE RANGE | TARGET |
|---|---|---|---|---|
| LEVEL I | I-70-19 | 10/2010 | 6.0 – 7.4 | 6.7 |
| LEVEL II | I-70-20 | 10/2010 | 12.1 – 14.9 | 13.5 |
| LEVEL III | I-70-18 | 10/2010 | 15.8 – 19.3 | 17.5 |

## TEST VALUES/RESULTS

| SERA ID | 1 | 2 | MEAN | # PASS |
|---|---|---|---|---|
| LEVEL I | 6.6 | 7.2 | 6.9 | 2/2 |
| LEVEL II | 13.1 | 13.0 | 13.1 | 2/2 |
| LEVEL III | 17.0 | 17.2 | 17.1 | 2/2 |

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 03/17/2010