



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **INORGANIC PHOSPHORUS REAGENT SET (UV METHOD)** MANUFACTURE DATE: 05/15/2014  
 CATALOG #: 1516-480  
 LOT #: 54929 EXP. DATE: 10/2016

<u>COMPONENTS:</u>	<u>CAT. #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. INORGANIC PHOSPHORUS UV REAGENT	I516-A	54660	11/2016
2. INORGANIC PHOSPHORUS UV STANDARD (5.0 mg/dL)	I516-B	54448	10/2016

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> <u>(on TC- 84)</u>	<u>TARGET</u> <u>(mg/dL)</u>
27611	05/2015	3.6 – 4.4	4.0
27602	03/2015	6.3 – 7.9	7.1

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	3.7	PASS
27602	7.1	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

Jacqueline Sanchez  
 Quality Control

DATE:

05/19/2014