



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **IRON / TIBC** MANUFACTURE DATE: 12/18/2009
 CATALOG #: I592-100G
 LOT #: 23159 EXP. DATE: 12/2011

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. IRON BUFFER REAGENT	I592-A	20973	12/2011
2. UIBC BUFFER REAGENT	I592-B	20974	12/2011
3. IRON COLOR REAGENT	I592-C	20975	12/2011
4. IRON STANDARD (500 ug/dL)	I592-D	20976	12/2011

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE		TARGET (ug/dL)	
		IRON	TIBC	IRON	TIBC
15261	08/2009	127 – 207	233 – 349	167	291
15262	08/2009	35 – 59	118 – 176	47	147

TEST VALUES/RESULTS

SERA ID	IRON		TIBC		MEAN		# PASS
	1	2	1	2	IRON	TIBC	
15261	161	146	296.6	260	153.5	265.1	4/4
15262	52	47	175.6	158.7	265.1	157.2	4/4

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 12/18/2009
 Quality Control