



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **IRON / TIBC** MANUFACTURE DATE: 05/19/2011
 CATALOG #: I592-100
 LOT #: 32639 EXP. DATE: 07/2013

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. IRON BUFFER REAGENT	I592-A	32423	07/2013
2. UIBC BUFFER REAGENT	I592-B	32424	07/2013
3. IRON COLOR REAGENT	I592-C	32427	07/2013
4. IRON STANDARD (500 ug/dL)	I592-D	32426	07/2013

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE		TARGET (ug/dL)	
		IRON	TIBC	IRON	TIBC
15281	09/2012	153-229	228-342	191	285
15282	09/2012	34-65	153-230	49	191

TEST VALUES/RESULTS

SERA ID	IRON		TIBC		MEAN		# PASS
	1	2	1	2	IRON	TIBC	
15281	200	204	332	338	202	335	4/4
15282	57	57	217	212	57	214.5	4/4

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 05/20/2011