



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **LDH – L LIQUID** MANUFACTURE DATE: 02/03/2010
CATALOG #: L535-60
LOT #: 22525 EXP. DATE: 05/2011

COMPONENTS: 1. LDH-L LIQUID REAGENT 1 LOT#: 17318 EXP. DATE: 05/2011

REAGENT BLANK VS. WATER: 0.342 AT 340 NM

TOLERANCE LEVELS OF SERUM CONTROL

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u>
15271	03/2012	115-173	144
15272	03/2012	273-409	341

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u>#PASS</u>
15271	151.1	151.1	151.1	2/2
15272	376.0	310.5	343.3	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Casanova* DATE: 01/03/2010
Quality Control