



**CERTIFICATE OF ANALYSIS
 (FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **MAGNESIUM** MANUFACTURE DATE: 04/14/2010
 CATALOG #: M527-100
 LOT #: 23227 EXP. DATE: 08/2011

COMPONENTS:	1. COLOR REAGENT	LOT#:	25070	EXP. DATE:	04/2013
	2. BUFFER REAGENT	LOT#:	16587	EXP. DATE:	08/2011
	3. STANDARD	LOT#:	20340	EXP. DATE:	08/2012

BLANK: 0.914 AT 550 NM

STANDARD: 2.0 mEq/L

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mEq/L)
15271	03/2012	1.0-1.7	1.35
15272	03/2012	2.7-4.2	3.45

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	1.3	1.5	1.4	2/2
15272	3.6	3.3	3.5	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
 BY:

Quality Control

DATE: 04/14/2010