



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **INORGANIC PHOSPHORUS REAGENT SET (UV METHOD)** MANUFACTURE DATE: 09/08/2011  
 CATALOG #: 1516-480  
 LOT #: 34730 EXP. DATE: 11/2013

<u>COMPONENTS:</u>	<u>CAT. #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. INORGANIC PHOSPHORUS UV REAGENT	I516-A	32771	11/2013
2. INORGANIC PHOSPHORUS UV STANDARD (5.0 mg/dL)	I516-B	32772	11/2013

BLANK: 0.062 AT 340 NM

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
15281	09/2012	2.6 – 3.9	3.2
15282	09/2012	5.3 – 7.9	6.6

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	3.2	3.3	3.3	2/2
15282	6.6	6.6	6.6	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED  
BY:

  
Quality Control

DATE: 09/08/2011